

# checkCIF/PLATON report

Structure factors have been supplied for datablock(s) dnba\_100

No syntax errors found.      CIF dictionary      Interpreting this report

## Datablock: dnba\_100

---

Bond precision:    C-C = 0.0015 A                      Wavelength=0.71073  
Cell:                      a=20.370(5)              b=8.760(5)              c=9.690(5)  
                                    alpha=90              beta=110.00(1)              gamma=90  
Temperature:              100 K

|                | Calculated  | Reported    |
|----------------|-------------|-------------|
| Volume         | 1624.8(13)  | 1624.8(13)  |
| Space group    | C 2/c       | C2/c        |
| Hall group     | -C 2yc      | ?           |
| Moiety formula | C7 H4 N2 O6 | ?           |
| Sum formula    | C7 H4 N2 O6 | C7 H4 N2 O6 |
| Mr             | 212.12      | 212.00      |
| Dx,g cm-3      | 1.734       | 1.733       |
| Z              | 8           | 8           |
| Mu (mm-1)      | 0.156       | 0.000       |
| F000           | 864.0       | 68.0        |
| F000'          | 864.63      |             |
| h,k,lmax       | 40,17,19    | 34,15,17    |
| Nref           | 6942        | 2725        |
| Tmin,Tmax      |             |             |
| Tmin'          |             |             |

Correction method= Not given

Data completeness= 0.393                      Theta(max)= 45.660

R(reflections)= 0.0388( 2143)              wR2(reflections)= 0.1198( 2725)

S = 0.866                      Npar= 173

---

The following ALERTS were generated. Each ALERT has the format

**test-name\_ALERT\_alert-type\_alert-level.**

Click on the hyperlinks for more details of the test.

---

### Alert level A

TYPE031\_ALERT\_1\_A \_diffrn\_radiation\_wavelength is not of type numb.  
PLAT027\_ALERT\_3\_A \_diffrn\_reflns\_theta\_full (too) Low ..... 0.00 Deg.  
PLAT029\_ALERT\_3\_A \_diffrn\_measured\_fraction\_theta\_full Low ..... 0.000  
PLAT091\_ALERT\_1\_A No Wavelength found in CIF - 0.71073 Ang Assumed ?

---

**Alert level B**

PLAT430\_ALERT\_2\_B Short Inter D...A Contact O5 .. N1 .. 2.80 Ang.

---

**Alert level C**

PLAT355\_ALERT\_3\_C Long O-H Bond (0.82A) O2 - H4 ... 1.01 Ang.

---

**Alert level G**

ABSMU\_01 Radiation type not identified. Calculation of  
\_exptl\_absorpt\_correction\_mu not performed.

PLAT005\_ALERT\_5\_G No \_iucr\_refine\_instructions\_details in CIF .... ?  
PLAT153\_ALERT\_1\_G The su's on the Cell Axes are Equal ..... 0.00500 Ang.  
PLAT180\_ALERT\_4\_G Check Cell Rounding: # of Values Ending with 0 = 3  
PLAT195\_ALERT\_1\_G Missing \_cell\_measurement\_theta\_max datum .... ?  
PLAT196\_ALERT\_1\_G Missing \_cell\_measurement\_theta\_min datum .... ?  
PLAT432\_ALERT\_2\_G Short Inter X...Y Contact O6 .. C7 .. 2.92 Ang.  
PLAT950\_ALERT\_5\_G Reported and Calculated Hmax Values Differ by .. 6  
PLAT951\_ALERT\_5\_G Reported and Calculated Kmax Values Differ by .. 2  
PLAT952\_ALERT\_5\_G Reported and Calculated Lmax Values Differ by .. 2  
PLAT981\_ALERT\_1\_G No non-zero f" Anomalous Scattering Values Found ?  
PLAT982\_ALERT\_1\_G The C-f' = 0.000 Deviates from the IT-value 0.003  
PLAT982\_ALERT\_1\_G The N-f' = 0.000 Deviates from the IT-value 0.006  
PLAT982\_ALERT\_1\_G The O-f' = 0.000 Deviates from the IT-value 0.011  
PLAT983\_ALERT\_1\_G The C-f" = 0.000 Deviates from the IT-Value 0.002  
PLAT983\_ALERT\_1\_G The N-f" = 0.000 Deviates from the IT-Value 0.003  
PLAT983\_ALERT\_1\_G The O-f" = 0.000 Deviates from the IT-Value 0.006

---

- 4 **ALERT level A** = Most likely a serious problem - resolve or explain  
1 **ALERT level B** = A potentially serious problem, consider carefully  
1 **ALERT level C** = Check. Ensure it is not caused by an omission or oversight  
16 **ALERT level G** = General information/check it is not something unexpected
- 12 ALERT type 1 CIF construction/syntax error, inconsistent or missing data  
2 ALERT type 2 Indicator that the structure model may be wrong or deficient  
3 ALERT type 3 Indicator that the structure quality may be low  
1 ALERT type 4 Improvement, methodology, query or suggestion  
4 ALERT type 5 Informative message, check
-

It is advisable to attempt to resolve as many as possible of the alerts in all categories. Often the minor alerts point to easily fixed oversights, errors and omissions in your CIF or refinement strategy, so attention to these fine details can be worthwhile. In order to resolve some of the more serious problems it may be necessary to carry out additional measurements or structure refinements. However, the purpose of your study may justify the reported deviations and the more serious of these should normally be commented upon in the discussion or experimental section of a paper or in the "special\_details" fields of the CIF. checkCIF was carefully designed to identify outliers and unusual parameters, but every test has its limitations and alerts that are not important in a particular case may appear. Conversely, the absence of alerts does not guarantee there are no aspects of the results needing attention. It is up to the individual to critically assess their own results and, if necessary, seek expert advice.

### **Publication of your CIF in IUCr journals**

A basic structural check has been run on your CIF. These basic checks will be run on all CIFs submitted for publication in IUCr journals (*Acta Crystallographica*, *Journal of Applied Crystallography*, *Journal of Synchrotron Radiation*); however, if you intend to submit to *Acta Crystallographica Section C* or *E*, you should make sure that full publication checks are run on the final version of your CIF prior to submission.

### **Publication of your CIF in other journals**

Please refer to the *Notes for Authors* of the relevant journal for any special instructions relating to CIF submission.

---

**PLATON version of 19/04/2012; check.def file version of 14/04/2012**

